


CERTIFICATE OF ANALYSIS No.: 2022-9843

CLIENT

Pharmahemp d.o.o., Cesta v Gorice 8
1000 Ljubljana, Slovenija

SAMPLE *

CBD DROPS PRM BLK 6,6% - olive oil










Sample condition: SUITABLE
Sample ID: 2237054
Sample type: Viscous liquid
Batch No.: * DR06622258B

Work order: 2022-106916
Analysis ID: 2022_208
Method ID: PHL_RPC_12C
Method SOP: MET-LAB-003-02

Sample received: 15/09/2022
Start of analysis: 15/09/2022
End of analysis: 16/09/2022
Analyst: Janez Gerdenc

* Information provided by the client.

CANNABINOID PROFILE	Concentration [% w/w]	Expanded uncertainty [% w/w]	Graphic presentation of relative cannabinoid concentration
CBDV - Cannabidivarin	0.734	0.088	
CBDA - Cannabidiolic acid	0.068	0.016	
CBGA - Cannabigerolic acid	< LOQ	n/a	
CBG - Cannabigerol	0.143	0.036	
CBD - Cannabidiol	6.50	0.32	
THCV - Tetrahydrocannavarin	0.256	0.041	
CBN - Cannabinol	< LOQ	n/a	
Δ⁹-THC - Δ-9-Tetrahydrocannabinol	< LOQ	n/a	
Δ⁸-THC - Δ-8-Tetrahydrocannabinol	< LOQ	n/a	
CBL - Cannabicyclol	< LOQ	n/a	
CBC - Cannabichromene	< LOQ	n/a	
Δ⁹-THCA - Δ-9-Tetrahydrocannabinolic acid	< LOQ	n/a	
CBE - Cannabielsoin	0.058 #	0.016	
CBNV - Cannabivarin	0.0389 #	0.0085	
CBCA - Cannabichromenic acid	< LOQ #	n/a	
CBT - Cannabicitran	< LOQ #	n/a	

Units and abbreviations: % w/w = weight percent, < LOQ = below the limit of quantitation (0.03 % w/w), ND = not detected, n/a = not available.

The results given herein apply only to the sample as received. **Expanded Uncertainty** was calculated using coverage factor k = 2, corresponding to a double standard uncertainty and characterizes the interval value in which it is possible to expect the real value with a probability of 95%. This is stated according to the ISO/IEC Guide 98-3.

Total or partial reproduction of this document is not allowed without the permit from PharmaHemp d.o.o. The document does not substitute any other legal document.

Date issued:

16/09/2022

Approved by:



mag. Marko Dragan
Analytical Laboratory Manager

Authorized by:



dr. Boštjan Jančar
Chief Technology Officer

End of Certificate